



Product Specifications Sheet

Specification Sheet No	SSN-PSW(NS)-002		
Product Name	Plain Swabs		
Product Description	Gauze Swabs		
Range	Premier		
Manufacturer	Mediq Healthcare UK Ltd		
Material	100% Cotton Gauze Type 13 Light		
Sterilisation	Non-sterile		
Issue Date	21st May 2024	Version No	R02



Intended use	Used during non-invasive procedures, absorbing blood and exudate, skin cleansing, wound padding and protection, secondary wound dressings, etc.
---------------------	---

Packaging						
Product code	Colour	Dimension (cm)	Ply	Items Per Pack	Qty- Inner	Qty – Shipper Carton
1650A	White	5 x 5	8	120	100	12,000
1665A	White	5 x 5	12	80	100	8,000
1655A	White	7.5 x 7.5	8	100	60	6,000
1670N	White	7.5 x 7.5	12	50	100	5,000
1660A	White	10 x 10	8	100	60	6,000
1675A	White	10 x 10	12	100	40	4,000
7011N	White	10 x 10	16	24	100	2,400
7102	White	10 x 10	32	24	50	1,200
7002A	Green	7.5 x 7.5	8	60	100	6,000
7003N	Green	7.5 x 7.5	12	N/A	100	5,000
7004N	Green	10 x 10	8	50	100	5,000
7005A	Green	10 x 10	12	40	100	4,000
7006A	Green	10 x 10	16	40	100	4,000
7007N	Green	10 x 10	32	N/A	50	1,200
7008N	Blue	10 x 10	8	50	100	5,000
7009A	Blue	10 x 10	12	40	100	4,000
7010N	Blue	10 x 10	32	24	50	1,200

Document Reference	Document Name	Reviewed By	Approved By	Revision Number	Issue Date
DMS 04-004	Product Specification Sheet	Jack Wong	Racheal Begg	R02	03 rd April 2024

Quality Certification	ISO 13485		
Device Classification	I	Rule	4
Conformity Route	Medical Devices Regulation (EU) 2017/745 Medical Devices Regulations 2002 (SI 2002 No 618, as amended) (UK MDR 2002)		
Storage conditions	Store in a cool dry place. Keep away from heat and direct sunlight.		
Shelf life	5 years from the date of manufacture with the recommended storage conditions		
Contraindications/Warnings	None		
Markings	Premier Logo, Code Ref, Lot, Product name, BP Type 13, Dimension, Ply, Colour, Single use, CE Marking, Latex free, Non-Sterile, Expiry date, Manufacturer's name and address, GSI Scan and barcode		
Additional information	100% Cotton Absorbent Gauze in compliance with EN 14079/BP 1988 standards. The swabs are soft, strong and highly absorbent.		
<i>Medical Device and Reporting System: Incidents caused by this device must be reported immediately to our offices.</i>			

The information provided within this Specification Sheet is correct as of last revision but it is subject to change without notice.

365 Healthcare	Mediq House
T: +44 (0) 1332 413 700	Arundel Avenue, Castle Donington
W: https://mediq.co.uk/	Derby, DE74 2HJ

365 Healthcare is a trading division of Mediq Healthcare UK Ltd
Registered in England No. 62537

Amendment Record

This procedure is reviewed to ensure its continuing relevance to the system and process that it describes. A record of contextual additions or omissions is given below. This revision history lists all changes to this document, the date of each revision, and the reason for each revision.

Revision Number	Revision Date	Reason for Revision	Updated by
R01	23/06/2023	Update to Mediq UK Brand replacing DSN 454	Racheal Begg
R02	21/05/2024	Update Mediq Healthcare UK address	Jack Wong
R03	03/02/2025	Update with 7003N & 7007N	Jack Wong

This document is uncontrolled when printed. The master document is controlled electronically, and document users are responsible for ensuring printed copies are valid prior to use to ensure it represents current practice within Mediq Healthcare UK Ltd.

Document Reference	Document Name	Reviewed By	Approved By	Revision Number	Issue Date
DMS 04-004	Product Specification Sheet	Jack Wong	Racheal Begg	R02	03 rd April 2024