



SANDRINGHAM ACTIVE MATTRESS



Product Code:

HAMSAND001

Description:

- The Sandringham is Harvest Healthcare's replacement system with cells and foam
- 18 cells made of TPU material reducing noise while the cells inflate and deflate, on 50mm of foam
- Pillow section for patient stability
- All cells are interchangeable for ease of repair and cleaning
- Light mattress allowing for the carer to easily move the mattress
- 4-way stretch welded PU cover which is vapour permeable and water resistant and has a Crib 5 fire rating
- Compatible with all of Harvest Healthcare's pumps allowing for a range of maximum patient weights

Mattress Dimensions:

2000 x 900 x 160mm

Weight:

9kg

Max. Patient Weight:

Harvest 1 Pump: 25 Stone (158 kg)
Harvest 250 Pump: 31 Stone (200kg)

Pressure Relief Risk:

High

Pump Technical Specification:

Please see separate pump specification sheet

Cleaning Guidance:

Cover should be wiped down with hot water and mild detergent. If heavily soiled, the cover may be machine washed up to 65°C and tumble dried. For disinfection use sodium hypochlorite solution max 1000ppm. Use no bleach, biological or phenolic-based cleaners. Dry completely before storing.

Warranty

In terms of Harvest Healthcare Guarantee and Warranties the Harvest Sandringham mattress is guaranteed for three years from date of purchase.

Exclusions:

1. Removable covers are guaranteed for one year
2. Damage to fabric covers through incorrect use and penetration by sharp instruments is excluded

Important Safety Notice

Harvest Healthcare Ltd supplies equipment and appliances to cover a range of medical conditions. Specific choice of equipment is the sole responsibility of the clinical practitioner or carer overseeing individual patient care. Harvest Healthcare Ltd supplies a range of products such as foam mattresses, active air mattresses, fall protection mats and bed rail protection bumpers. These products are usually covered in vapour permeable PU fabric that could cause risk of suffocation in the event of prolonged facial contact. Mattress risk categories are to be used as a guide only. Risk assessments specific to user must be carried out prior to the use of all products. Products and data continually undergo development and may therefore deviate from the details given.

Version 2: March 2018

Unit 6/7, Block 4, Ashbourne Business Park, Co Meath t: 01 835 2411 e: orders@medguard.ie